

Ex.nr. 1

Obiectivul de investiții:

«Sondă pentru măsurarea câmpului electric în sistemul de
măsurare SAR »

Obiectivul supus recepției:

« Sondă pentru măsurarea câmpului electric în
sistemul de măsurare SAR (cod CPV:31711120-0),
furnizate de către S.C. Global Logistic Systems S.R.L.,
în baza Contractului nr. e-SC-5772/23.07.2024»

Concluzia: ADMIS

PROCES VERBAL DE RECEPȚIE
(Cod 14-2-5) cod angajament **AAATBTP6AF6**

I. DATE GENERALE

Comisia de recepție numită în baza notei de desemnare a comisiei de recepție nr. e-LICETER/162/30.09.2024, convocată la data de 06.11.2024 și-a desfășurat activitatea în data de 06.11.2024.

II. CONSTATĂRI

În urma examinării documentației prezentate și a cercetării pe teren a produselor furnizate, s-a constatat că produsul reprezentând „**Sondă pentru măsurarea câmpului electric în sistemul de măsurare SAR**”, cod CPV:31711120-0, furnizat de către S.C. Global Logistic Systems S.R.L., în baza Contractului nr. e-SC-5772/23.07.2024, este funcțional și are caracteristicile în conformitate cu prevederile din Caietul de sarcini nr. e-LICETER/24/12.03.2024, anexă la comandă.

Valoarea totală a produselor furnizate supuse recepției, conform documentelor de decontare este de: **101.974,00 Lei** [fără TVA], respectiv **121.349,06 Lei** [inclusiv TVA].

În cadrul recepției au fost întocmite/primate următoarele documente:


- aviz de însoțire marfă;
- certificat de calitate și garanție;
- declarație de conformitate/certificat de conformitate;
- certificat de etalonare (calibrare).

III. CONCLUZII

1. Pe baza constatărilor și concluziilor consemnate mai sus, comisia de recepție hotărăște în unanimitate:

**SE ADMITE RECEPȚIA OBIECTIVULUI DE INVESTIȚII:
*Sondă pentru măsurarea câmpului electric în sistemul de măsurare SAR (cod
CPV:31711120-0), ȘI SE ACORDĂ CALIFICATIVUL „FOARTE BINE”.***

2. În conformitate cu H.G. 2.139/2004 pentru aprobarea Catalogului privind clasificarea și duratele normale de funcționare a mijloacelor fixe, pentru produsul recepționat s-a stabilit durata normală de funcționare (vezi tabel).
3. Prezentul proces verbal, care conține 2 (două) file și o anexă, însumând 3 (trei) file a fost încheiat azi, 06.11.2024 în trei exemplare originale.




ANEXĂ LA PROCESUL VERBAL DE RECEPȚIE (cod 14-2-5)

nr. 124 din 06.11.2024 a produselor, în vederea extinderii capacității de măsurare a echipamentelor existente în cadrul LICETER, respectiv **Sondă pentru măsurarea câmpului electric în sistemul de măsurare SAR (cod CPV:31711120-0)** furnizată de către **S.C. Global Logistic Systems S.R.L.**, în baza Contractului nr. e-SC-5772/23.07.2024.

Nr. crt.	Denumire echipament	Cantitate	Preț unitar, fără TVA (lei)	Preț total, fără TVA (lei)	codul de clasificare cf. H.G. nr. 2139/2004	durata normală de funcționare cf. H.G. nr. 2139/2004 [ani]
1.	Sondă pentru măsurarea câmpului electric în sistemul de măsurare SAR	1	101.974,00	101.974,00	2.2.3.1	4
Valoarea totală a produselor, fără TVA [lei]:				101.974,00		
Valoarea totală a produselor, cu TVA [lei]:				121.349,06		

*) Prețul include toate costurile ofertantului, directe și indirecte, legate de încheierea și executarea contractului, în conformitate cu cerințele minime obligatorii prevăzute în Secțiunea 2: Caietul de sarcini din Documentația de atribuire.

ANCOM	Numele și prenumele	Semnătura



AVIZ DE INSOTIRE A MARFII

Serie GLS Numar 75

Data 04.11.2024

- RON -

Furnizor

Client

S.C.GLOBAL LOGISTIC SYSTEMS SRL

AUTORITATEA NATIONALA PENTRU ADMINISTRARE SI
REGLEMENTARE IN COM

CIF RO13930048 RC J40/5444/2001

CIF 14751237

Capital soc. 2 000 RON
BUCURESTI sect. 3 str. DIMITRIE BOLINTINEANU nr. 5 ap.
44
Telefon 0212527681 Email contact@glis.ro

BUCURESTI SECTOR 3, STR. DELEA NOUA, NR. 2

Banca ING BANK, IBAN (RON)
RO96INGB0000999907340337

Banca TREZORERIA STATULUI, IBAN(RON)
RO54TREZ7035069XXX001141

Nr. crt.	Denumire produse/servicii	UM	Cantitate	Pret unitar	Valoare	TVA (19%)
1	SONDA PENTRU MASURAREA CAMPULUI ELECTRIC IN SISTEM DE MASURARE SAR; COD PRODUS 3724-EPGO-411, SAR PROBE	BUC	1.000	101 974.0000	101 974.00	19 375.06
					101 974.00	19 375.06
Total						121 349.06





S.C. Global Logistic Systems SRL

Punct de lucru: Str. CORBENI, nr. 10, Sector 2, Bucuresti
J40/5444/2001, RO 13930048

IBAN: RO96INGB0000999907340337, ING BANK NV

IBAN: RO54TREZ7035069XXX001141, Trezoreria Sector 3

Tel. 021 252 76 81, Mobil (+4) 0722 661 266

Website: <http://www.gls.ro> ; e-mail: contact@glis.ro

CERTIFICAT DE CALITATE SI GARANTIE

Data 06.11.2024

Conform factura fiscala: seria GLS nr 17775/ 04.11.2024

Termenul de garantie pentru: Sonda pentru masurarea campului electric in sistem de masura SAR; respectiv cod produs: 3724-EPGO-411 SAR PROBE, este de 12 luni.

Declaram ca produsul este conform specificatiilor din oferta.

Nu fac obiectul garantiei, accesoriile rupte, sparturi, , zgarieturi, componente desprinse, defecte datorate scunfundarii in lichide sau manevrarii gresite ale produselor


Instructiunile de utilizare, depozitare, instalare, sunt furnizate odata cu echipamentele.

Instalarea sau interventia asupra echipamentelor si/sau accesoriilor de catre alte persoane decat cele autorizate de catre furnizor atrage pierderea garantiei.

Perioada de garantie se prelungeste cu durata cat echipamentele s-au aflat in reparatii.


Furnizorul poate decide daca in perioada de garantie va repara echipamentele defecte sau le va inlocui definitiv cu echipamente similare.

Furnizorul nu este responsabil pentru eventualele pierderi suferite de catre cumparator ca urmare a nefunctionarii echipamentelor.



DECLARATIE DE CONFORMITATE

Declaram pe propria raspundere ca echipamentele furnizate sunt conform cu specificatiile fabricantului.





COMOSAR E-Field Probe Calibration Report

Ref : ACR.275.31.24.BES.B

Cancel and replace the report ACR.275.31.24.BES.A

ANCOM
LOCALITATEA PREJMER, STRADA MITROPOLIT ANDREI
ŞAGUNA NR. 1033, JUDEŢUL BRAŞOV
MVG COMOSAR DOSIMETRIC E-FIELD PROBE
SERIAL NO.: 3724-EPGO-441

Calibrated at MVG
Z.I. de la pointe du diable
Technopôle Brest Iroise – 295 avenue Alexis de Rochon
29280 PLOUZANE - FRANCE

Calibration date: 10/10/2024



Accreditations #2-6789
Scope available on www.cofrac.fr

The use of the Cofrac brand and the accreditation references is prohibited from any reproduction.

Summary:

This document presents the method and results from an accredited COMOSAR Dosimetric E-Field Probe calibration performed at MVG, using the CALIPROBE test bench, for use with a MVG COMOSAR system only. The test results covered by accreditation are traceable to the International System of Units (SI).



COMOSAR E-FIELD PROBE CALIBRATION REPORT

Ref: ACR.275.31.24.BES.B

	<i>Name</i>	<i>Function</i>	<i>Date</i>	<i>Signature</i>
<i>Prepared by :</i>				
<i>Checked & approved by:</i>				
<i>Authorized by:</i>				

	<i>Customer Name</i>
<i>Distribution :</i>	ANCOM

<i>Issue</i>	<i>Name</i>	<i>Date</i>	<i>Modifications</i>
A			
B			



TABLE OF CONTENTS

1 Device Under Test 4

2 Product Description 4

 2.1 General Information 4

3 Measurement Method 4

 3.1 Sensitivity 4

 3.2 Linearity 5

 3.3 Isotropy 5

 3.4 Boundary Effect 5

 3.5 Probe Modulation Response 6

4 Measurement Uncertainty 6

5 Calibration Results..... 6

 5.1 Calibration in air 6

 5.2 Calibration in liquid 7

6 Verification Results 8

7 List of Equipment 9



1 DEVICE UNDER TEST

Device Under Test	
Device Type	COMOSAR DOSIMETRIC E FIELD PROBE
Manufacturer	MVG
Model	SSE2
Serial Number	3724-EPGO-441
Product Condition (new / used)	New
Frequency Range of Probe	0.15 GHz-7.5GHz
Resistance of Three Dipoles at Connector	Dipole 1: R1=0.227 MΩ Dipole 2: R2=0.232 MΩ Dipole 3: R3=0.223 MΩ

2 PRODUCT DESCRIPTION

2.1 GENERAL INFORMATION

MVG's COMOSAR E field Probes are built in accordance to the IEC/IEEE 62209-1528 and FCC KDB865664 D01 standards.



Figure 1 – MVG COMOSAR Dosimetric E field Probe

Probe Length	330 mm
Length of Individual Dipoles	2 mm
Maximum external diameter	8 mm
Probe Tip External Diameter	2.5 mm
Distance between dipoles / probe extremity	1 mm

3 MEASUREMENT METHOD

The IEC/IEEE 62209-1528 and FCC KDB865664 D01 standards provide recommended practices for the probe calibrations, including the performance characteristics of interest and methods by which to assess their effect. All calibrations / measurements performed meet the fore-mentioned standards.

3.1 SENSITIVITY

The sensitivity factors of the three dipoles were determined using a two step calibration method (air and tissue simulating liquid) using waveguides as outlined in the standards for frequency range 600-7500MHz and using the calorimeter cell method (transfer method) as outlined in the standards for frequency 150-450 MHz.



3.2 LINEARITY

The evaluation of the linearity was done in free space using the waveguide, performing a power sweep to cover the SAR range 0.01W/kg to 100W/kg.

3.3 ISOTROPY

The axial isotropy was evaluated by exposing the probe to a reference wave from a standard dipole with the dipole mounted under the flat phantom in the test configuration suggested for system validations and checks. The probe was rotated along its main axis from 0 to 360 degrees in 15-degree steps. The hemispherical isotropy is determined by inserting the probe in a thin plastic box filled with tissue-equivalent liquid, with the plastic box illuminated with the fields from a half wave dipole. The dipole is rotated about its axis (0°–180°) in 15° increments. At each step the probe is rotated about its axis (0°–360°).

3.4 BOUNDARY EFFECT

The boundary effect is defined as the deviation between the SAR measured data and the expected exponential decay in the liquid when the probe is oriented normal to the interface. To evaluate this effect, the liquid filled flat phantom is exposed to fields from either a reference dipole or waveguide. With the probe normal to the phantom surface, the peak spatial average SAR is measured and compared to the analytical value at the surface.

The boundary effect uncertainty can be estimated according to the following uncertainty approximation formula based on linear and exponential extrapolations between the surface and $d_{be} + d_{step}$ along lines that are approximately normal to the surface:

$$SAR_{uncertainty} [\%] = \Delta SAR_{be} \frac{(d_{be} + d_{step})^2}{2d_{step}} \frac{(e^{-\alpha_{be}(\delta/2)})}{\delta/2} \quad \text{for } (d_{be} + d_{step}) < 10 \text{ mm}$$

where

$SAR_{uncertainty}$	is the uncertainty in percent of the probe boundary effect
d_{be}	is the distance between the surface and the closest <i>zoom-scan</i> measurement point, in millimetre
Δ_{step}	is the separation distance between the first and second measurement points that are closest to the phantom surface, in millimetre, assuming the boundary effect at the second location is negligible
δ	is the minimum penetration depth in millimetres of the head tissue-equivalent liquids defined in this standard, i.e., $\delta \approx 14$ mm at 3 GHz;
ΔSAR_{be}	in percent of SAR is the deviation between the measured SAR value, at the distance d_{be} from the boundary, and the analytical SAR value.

The measured worst case boundary effect SARuncertainty[%] for scanning distances larger than 4mm is 1.0% Limit ,2%).



3.5 PROBE MODULATION RESPONSE

MVG’s probe were evaluated experimentally with various modulated signal and the deviation from CW response were found neglectable in the used power range of the probe. So the correction to taking into account the linearization parameters for different modulation is null, therefore the CW factor given in this report can be used whatever the measured modulation

4 MEASUREMENT UNCERTAINTY

The guidelines outlined in the IEC/IEEE 62209-1528 and FCC KDB865664 D01 standards were followed to generate the measurement uncertainty associated with a SAR probe calibration using the waveguide or calorimetric cell technique depending on the frequency.

The estimated expanded uncertainty (k=2) in calibration for SAR (W/kg) is +/-11% for the frequency range 150-450MHz.


The estimated expanded uncertainty (k=2) in calibration for SAR (W/kg) is +/-14% for the frequency range 600-7500MHz.

5 CALIBRATION RESULTS

Ambient condition	
Liquid Temperature	20 +/- 1 °C
Lab Temperature	20 +/- 1 °C
Lab Humidity	30-70 %

5.1 CALIBRATION IN AIR

The following curve represents the measurement in waveguide of the voltage picked up by the probe toward the E-field generated inside the waveguide.



From this curve, the sensitivity in air is calculated using the below formula.

$$E^2 = \sum_{i=1}^3 \frac{V_i (1 + V_i / DCP_i)}{Norm_i}$$



where

V_i =voltage readings on the 3 channels of the probe

DCPi=diode compression point given below for the 3 channels of the probe

Normi=dipole sensitivity given below for the 3 channels of the probe

Normx dipole 1 ($\mu\text{V}/(\text{V}/\text{m})^2$)	Normy dipole 2 ($\mu\text{V}/(\text{V}/\text{m})^2$)	Normz dipole 3 ($\mu\text{V}/(\text{V}/\text{m})^2$)
0.66	0.76	0.67

DCP dipole 1 (mV)	DCP dipole 2 (mV)	DCP dipole 3 (mV)
122	110	131

5.2 CALIBRATION IN LIQUID

The calorimeter cell or the waveguide is used to determine the calibration in liquid using the formula below.

$$\text{Conv}F = \frac{E_{\text{liquid}}^2}{E_{\text{air}}^2}$$

The E-field in the liquid is determined from the SAR measurement according to the below formula.

$$E_{\text{liquid}}^2 = \frac{\rho \text{ SAR}}{\sigma}$$

where

σ =the conductivity of the liquid

ρ =the volumetric density of the liquid

SAR=the SAR measured from the formula that depends on the setup used. The SAR formulas are given below

For the calorimeter cell (150-450 MHz), the formula is:

$$\text{SAR} = c \frac{dT}{dt}$$

where

c =the specific heat for the liquid

dT/dt =the temperature rises over the time

For the waveguide setup (600-75000 MHz), the formula is:

$$\text{SAR} = \frac{4P_W}{ab\delta} e^{-\frac{2z}{\delta}}$$

Page: 7/10





where

- a=the larger cross-sectional of the waveguide
- b=the smaller cross-sectional of the waveguide
- δ =the skin depth for the liquid in the waveguide
- Pw=the power delivered to the liquid

The below table summarize the ConvF for the calibrated liquid. The curves give examples for the measured SAR depending on the voltage in some liquid.


Liquid	Frequency (MHz*)	ConvF
HL750	750	2.81
HL835	835	2.68
HL900	900	2.71
HL1800	1800	2.70
HL1900	1900	2.91
HL2000	2000	3.16
HL2450	2450	3.07
HL2600	2600	2.93
HL3500	3500	2.50
HL3700	3700	2.44

(*) Frequency validity is +/-50MHz below 600MHz, +/-100MHz from 600MHz to 6GHz and +/-700MHz above 6GHz




6 VERIFICATION RESULTS

The figures below represent the measured linearity and axial isotropy for this probe. The probe specification is +/-0.2 dB for linearity and +/-0.15 dB for axial isotropy.



Linearity: +/-1.91% (+/-0.08dB)



Isotropy: +/-0.14% (+/-0.01dB)



7 LIST OF EQUIPMENT

Equipment Summary Sheet				
Equipment Description	Manufacturer / Model	Identification No.	Current Calibration Date	Next Calibration Date
CALIPROBE Test Bench	Version 2	NA	Validated. No cal required.	Validated. No cal required.
Network Analyzer	Rohde & Schwarz ZVM	100203	08/2021	08/2026
Network Analyzer – Calibration kit	Rohde & Schwarz ZV-Z235	101223	07/2022	07/2025
Multimeter	Keithley 2000	4013982	02/2023	02/2026
Signal Generator	Rohde & Schwarz SMB	106589	03/2022	03/2025
Amplifier	MVG	MODU-023-C-0002	Characterized prior to test. No cal required.	Characterized prior to test. No cal required.
Power Meter	NI-USB 5680	170100013	06/2021	06/2026
USB Sensor	Keysight U2000A	SN: MY62340002	10/2022	10/2025
Directional Coupler	Krytar 158020	131467	Characterized prior to test. No cal required.	Characterized prior to test. No cal required.
Fluoroptic Thermometer	LumaSense Luxtron 812	94264	09/2022	09/2025
Coaxial cell	MVG	SN 32/16 COAXCELL_1	Validated. No cal required.	Validated. No cal required.
Waveguide	MVG	SN 32/16 WG2_1	Validated. No cal required.	Validated. No cal required.
Liquid transition	MVG	SN 32/16 WGLIQ_0G600_1	Validated. No cal required.	Validated. No cal required.
Waveguide	MVG	SN 32/16 WG4_1	Validated. No cal required.	Validated. No cal required.
Liquid transition	MVG	SN 32/16 WGLIQ_0G900_1	Validated. No cal required.	Validated. No cal required.
Waveguide	MVG	SN 32/16 WG6_1	Validated. No cal required.	Validated. No cal required.
Liquid transition	MVG	SN 32/16 WGLIQ_1G500_1	Validated. No cal required.	Validated. No cal required.
Waveguide	MVG	SN 32/16 WG8_1	Validated. No cal required.	Validated. No cal required.
Liquid transition	MVG	SN 32/16 WGLIQ_1G800B_1	Validated. No cal required.	Validated. No cal required.
Liquid transition	MVG	SN 32/16 WGLIQ_1G800H_1	Validated. No cal required.	Validated. No cal required.



COMOSAR E-FIELD PROBE CALIBRATION REPORT

Ref: ACR.275.31.24.BES.B

Waveguide	MVG	SN 32/16 WG10_1	Validated. No cal required.	Validated. No cal required.
Liquid transition	MVG	SN 32/16 WGLIQ_3G500_1	Validated. No cal required.	Validated. No cal required.
Waveguide	MVG	SN 32/16 WG12_1	Validated. No cal required.	Validated. No cal required.
Liquid transition	MVG	SN 32/16 WGLIQ_5G000_1	Validated. No cal required.	Validated. No cal required.
Waveguide	MVG	SN 32/16 WG14_1	Validated. No cal required.	Validated. No cal required.
Liquid transition	MVG	SN 32/16 WGLIQ_7G000_1	Validated. No cal required.	Validated. No cal required.
Temperature / Humidity Sensor	Testo 184 H1	44235403	02/2024	02/2027

Tip Entitate Entitate Publica Alte Entitati Persoana Fizica

ORDIN DE PLATA MULTIPLU ELECTRONIC (OPME)

Suma control 15029071

Numar document 0000000846

Data document 18.11.2024

Nr.inregistrari 2

F1129 A2.0.31

Total Document 277.834,06 LEI LEI ADICA douasutesaptezeceisaptemilioptsutetrezecisipatrulei.06bani

Denumire Plator ANCOM

CIF plator 14751237

Adresa plator DELEA NOUA 2 SECTOR 3

PLATITOR

BENEFICIAR

NR. OP.	Cod IBAN plator	Denumire Trezorerie	Cod Program	Cod Angajament	Indicator Angaj.	Denumire Beneficiar	Cif/CNP Beneficiar	Cod IBAN beneficiar	Denumire Banca / Trezorerie	Suma	Numar de evidenta a platii	Reprezentand
564T	RO12TREZ23F850100710102X	TREZORERIA STATULUI	0000000000	AAATBTP6AF6	AAB	GLOBAL LOGISTIC SYSTEMS SRL	13930048	RO54TREZ70350690XX001141	Trezorerie Sector 3	121.349,06		FFGLS 17775 04112024 PVR NR LICETI 124 06112024 CTR eSC 5772 230724

VALIDARE SI GENERARE XML

Semnatura 1

Semnatura 2

FISIERELE CARE SE IMPORTA TREBUIE SA AIBA TERMINATIA .XML, SA FI FOST GENERATE DIN VERSIUNI ANTERIOARE ALE ACESTUI FORMULAR SAU SA FI FOST GENERATE DIN APLICATIILE BENEFICIARULUI, CU RESPECTAREA INTOCMAI A STRUCTURII DIN GHIDUL UTILIZATORULUI.

NERESPECTAREA ACESTOR INSTRUCIUNI VA PROVOCA O FUNCTIONARE DEFECTUOASA A FORMULARULUI.

ATENTIE!!! Debitarea contului se va realiza in functie de data si ora la care se depune documentul si se primeste prima notificare din partea sistemului prin care se confirma corectitudinea formala a documentului.

Export XML

Import XML